Bowie Dick Test Pack

SAP / Monitoring

Once a day ..... since 1963

REF: 102.001.0020 Unit: 20 pieces

Used for daily monitoring of pre-vacuum steam sterilizer. Based upon the Bowie Dick test result the steam sterilizer is released for usage during the day.

Characteristics:

- ✦ Fulfill requirements of EN 285 and ISO 17665
- ✦ Conform to ISO 11140 - 1 class 2
- ✦ Independently tested to show performance similar to original pack
- ✦ Clear yellow - black color change
- ✦ Easy to interpret
- ✦ Easy to document
- ✦ Non-Toxic / Lead Free

The Bowie Dick test detect failures like air leaks, insufficient vacuum, inadequate steam penetration.

Majac Medical Products Pty Ltd
Ph: +617 3265 6355
Fax: +617 3865 2729
sales@majacmedical.com.au
www.majacmedical.com.au

FS 540017

Also Available From: